

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

NOVEMBER 25, 2009

OPERATOR'S NAME:

RITALA

OPERATOR'S NUMBER: 3791

SUBJECT'S LAST NAME:

VERIFICATION OF CALIBRATION

SUBJECT'S FIRST NAME/MI :

INSTALL NEW ALCO TANK

O.L. #: A

DEPT/AGENCY: ANC1

CASE/REPORT: A

TEST TYPE: V

ALCO TARGET VALUE: 081

ALCO S/N: 78895

### --- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.31 in		
ALCO TARGET	.079	16:08
BLANK TEST	.000	16:09
INTERNAL STANDARD	VERIFIED	16:09
ALCO TV 29.31 in	.081	16:09
BLANK TEST	.000	16:10
SUBJECT SAMPLE	.000	16:10
BLANK TEST	.000	16:11
ALCO TV 29.31 in	.081	16:12
BLANK TEST	.000	16:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130579

NOVEMBER 25, 2009

TIME 16:13

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	12/11/08
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	29.31 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~Δ

# VERIFICATION OF CALIBRATION REPORT DEC 03 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130579
Name		Chris RITAN	ID# 3791 Date 11/25/09
A	Agency	ANCHORAGE POLICE DEPT.	Phone # 786-2640
Instrument Location		ANCHORAGE JAIL	
Alco S/N		78895	Target Value 081 High Pressure 1050
B	Alco Test Value Average	081	081
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		Chris RITAN / 1349	
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of DEC, 2009.

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

